

Application Note Glycans & Glycoprotein



Analysis of N-Glycans using HPAEC-PAD/MS

Introduction

Glycosylation is a fundamental biological process in which polysaccharides, referred to as glycans, become covalently bound to proteins. In the field of glycomics, one of the major challenges is the multitude of distinct glycan isomers found within a glycoprotein, each varying in their abundances [1]. To improve the structure elucidation workflow in glycomics, the development of analytical methods to detect and quantify glycans is crucial. Many widely used analytical techniques for glycans detection and quantification involve the derivatization of glycan samples after their release from glycoprotein. However, the derivatization may cause some issues such as incomplete labelling or loss of sialic acids, ultimately leading to incomplete structure elucidation [2]. Therefore, a glycan analysis method that eliminates the need for derivatization is preferrable.

High-performance anion exchange chromatography combined with pulsed amperometric detection (HPAEC-PAD) coupled with mass spectroscopy (MS) detection offers an optimal analytical approach for sensitive detection of *N*-glycans without sample derivatization. Previous studies have highlighted the use of HPAEC-PAD [3-4] or the use of a mass spectrometer [5-6] for the detection of *N*-glycans. Notably, several studies have combined mass spectrometry for glycan identification with PAD for glycan quantification [7-9]. In this application note, a glycan detection method is presented based on the HPAEC-PAD/MS technique. A sialylated *N*-glycans standard from fetuin was analyzed as a proof of principle to demonstrate the performance of the method.

Method

Separation of the N-glycans is based on HPAEC using the new SweetSep[™] AEX200 anion-exchange column. Carbohydrates can be separated using HPAEC under alkaline conditions (pH > 12). N-glycans are polysaccharides, which under alkaline conditions will be retained stronger than simple monosaccharides or disaccharides. Therefore, gradient elution with an increasing amount of sodium acetate modifier is used to speed up the elution of the N-glycans (see gradient program in Table 1). Subsequently, a column clean-up step (100 mM NaOH + 200 mM NaOAc) is executed for 5 minutes, followed by 15 minutes of equilibration to the starting conditions. It is essential to perform a desalting step after the separation on the column. Sodium hydroxide and acetate eluents are non-volatile and high conductance solvents and are not compatible with



Figure 1. Instrument setup for N-glycan analysis using HPAEC-PAD/MS.

Electrochemistry Discover the difference

Table 1. HPAEC-PAD conditions

| LC system | Quaternary HPLC system |
|--------------------------------------|--|
| Detector | DECADE Elite electrochemical detector |
| Columns | SweetSep™ AEX200, 4 × 200 mm column, 5 μm |
| Flow rate | 0.7 mL/min and 1 mL/min |
| Gradient program | 0 min: 100 mM NaOH + 6 mM NaOAc |
| | 70 min: 100 mM NaOH + 190 mM NaOAc |
| | 70—75 min: 100 mM NaOH + 200 mM NaOAc |
| | 75—90 min: 100 mM NaOH + 6 mM NaOAc |
| System backpressure | About 310 bar |
| Temperature | 30°C for separation, 35°C for detection |
| Flow cell | FlexCell Au WE, HyREF, carbon filled PTFE AUX |
| PAD Potential wave- form (4-step) | E1, E2, E3, E4: +0.1, -2.0, +0.6, -0.1 V ts. t1, t2, t3, t4: 0.2, 0.4, 0.02, 0.01, 0.07 s |
| ····· (· •···p) | |

Table 2. Desalting & MS conditions

| Detector | Bruker Daltonics HCT Plus (ESI-ion trap) |
|------------------------|--|
| Data acquisition | Bruker Compass & EsquireControl software |
| Desalter | Thermo Scientific™ Dionex™ ERD™ 500 Electrolytically Regenerated Desalter |
| Desalter current | 500 mA |
| Desalter potential | 4.0 V |
| Capillary potential | 4000 V |
| End plate potential | 3500 V |
| Nebulizer pressure | 60 psi |
| Drying gas flow rate | 10 L/min |
| Drying gas temperature | 365°C |

electrospray ionization MS and may lead to damage of the ion source. Moreover, the presence of non-volatile salts can lead to ion suppression and thus a loss in sensitivity of MS detection [10]. The use of the desalter can also be circumvented using volatile buffer system such as ammonium formate [10, 11].

Prior to the detection, the effluent from the desalter is split with a 1:1 split ratio to allow simultaneous PAD & MS detection (Figure 1). The HPAEC-PAD conditions are listed in Table 1. For PAD detection, a 4-step potential waveform was applied. The choice of the 4-step potential waveform resulted in excellent reproducibility and minimal electrode wear [12]. The detection temperature was set at 35°C. MS and desalting conditions are listed in Table 2.

Results

An overlay of chromatograms of 10 μ L injections of 25 μ M fetuin oligosaccharide standard is shown in Figure 2. The overlay consists of chromatograms obtained from the pulsed amperometric detection and the MS detection (represented by the Total Ion Current (TIC) chromatogram). The assignment of peaks was done by comparative analysis of the acquired chromatograms and existing literature references [13, 14], as



Figure 2. Chromatograms obtained with an 10 μ L injection of a 25 μ M N-glycans standard containing di– and tri–sialylated oligosaccharides on a SweetSep[™] AEX 200 column, 4.0 mm ID × 200 mm. Top: Total Ion Current (TIC) Chromatogram. Bottom: Pulsed Amperometric Detection Chromatogram.



Figure 3. PAD Chromatograms obtained with an 10 μ L injection of a 25 μ M N-glycans standard showing the separation between the di-, tri-, and tetrasialylated N-glycans.



Figure 4. TIC Chromatograms obtained with an 10 μ L injection of a 25 μ M N-glycans standard on 0.7 mL/min flow rate.



the individual standards were not available. The oligosaccharide standard consists of a mixture of di-, tri-, and tetrasialylated *N*-glycans. Both PAD and TIC signals show consistent data corresponding to the disialylated and trisialylated *N*glycans. However, tetrasialylated *N*-glycans are only detected in the PAD mode under these conditions (Figure 3). This issue arises from the limited capacity of the desalter, resulting in suppressed ion signals caused by the presence of salts [10].

The presence of salts can be avoided using the pre-desalter flow splitting. The pre-desalter flow splitting is a popular option in several references because desalting is only needed for MS detection [7-9]. However, when implementing pre-desalting flow splitting, it is important to note that the flow rate of the effluent after the desalter is significantly decreased. Consequently, the reduced amount of N-glycans going into the MS could lead to a decrease in the signal-to-noise ratio and a potential loss of resolution. Alternatively, salt contamination on MS can be avoided by decreasing the concentration of the salt or the flow rate. Although effective in minimizing salt contamination, this option requires a longer run time to separate N-glycans. A TIC chromatogram in Figure 4 demonstrates that N-glycans were retained longer when the flow rate is decreased to 0.7 mL/min. However, the detection of tetrasialylated N-glycans is now enabled and in addition, the peak response is higher compared to the higher flow rate configuration.

In conclusion, the presented method shows the potential of parallel PAD and MS detection for intact *N*-glycan analysis. Further, this method can be extended into tandem MS/MS for the identification of *N*-glycans coupled with PAD for the quantification of *N*-glycans.

References

- Spiro, R.G., Protein glycosylation: nature, distribution, enzymatic formation, and disease implications of glycopeptide bonds, *Glycobiology*, 2002, 12(4), 43R-56R
- Szabo, Z., Thayer, J.R., Rohrer, J., Instrument configuration for native *N*-linked oligosaccharide characterization by HPAE-PAD/MS, ThermoFisher Scientific Technical Note 72478
- Townsend, R.R., et al., Separation of branched sialylated oligosaccharides using high-pH HPAEC-PAD, *Anal. Biochem.*, 1989, 182, 1-8
- Rohrer, J.S., Separation of asparagine-linked oligosaccharides by high-pH anion-exchange chromatography with pulsed amperometric detection: empirical relationships between oligosaccharide structure and chromatographic retention, *Glycobiology*, 1995, 5(4), 359-363
- 5. Han, L., Costello, C.E., Mass spectrometry of glycans, *Biochemistry*, 2013, 78(7), 710-720

For research purpose only. The information shown in this communication is solely to demonstrate the applicability of the ALEXYS system. The actual performance may be affected by factors beyond Antec's control. Optimization of the method and sample preparation may be necessary for analysis of specific real samples. Specifications mentioned in this application note are subject to change without further notice.

- Liew, C.Y., et al., Structural identification of N-glycan isomers using logically derived sequence tandem mass spectrometry, Commun. Chem., 2021, 4:92
- Maier, M., et al., Applying mini-bore HPAEC-MS/MS for the characterization and quantification of Fc N-glycans from heterogeneously glycosylated IgGs, *J. Chrom. B.*, 2016, 1033-1034, 342-352
- Kandzia, S., Costa, J., N-glycosylation analysis by HPAEC-PAD and mass spectrometry, *Methods Mol. Biol.*, 2013, 1049, 301-312
- Bruggink, C., et al., Analysis of carbohydrates by anion exchange chromatography and mass spectrometry, *J. Chrom. A.*, 2005, 1085, 104-109
- 10. Flick, T.G., et al., Solution additives that desalt protein ions in native mass spectrometry, *Anal. Chem.*, 2012, 84, 7511-7517
- van Schaick, G., et al., Anion exchange chromatography mass spectrometry for monitoring multiple quality attributes of erythropoietin biopharmaceuticals, *Anal. Chim. Acta*, 2021, 1143, 166-172
- Rocklin, R.D., Clarke, A.P., Weitzhandler, M., Improved longterm reproducibility for pulsed amperometric detection of carbohydrates via a new quadruple-potential waveform, *Anal. Chem*, 1998, 70, 1496-1501
- Baenziger, J.U., Fiete, D., Structure of the complex oligosaccharides of fetuin, *J. Biol. Chem.*, 1979, 254(3), 789-795
- 14. Certificate of Analysis Lot Number 190508 (Expired on 2029), ThermoFisher Scientific, 1997, Doc. No. 034447-20

Antec Scientific (USA) info@AntecScientific.com www.AntecScientific.com T 888 572 0012

Antec Scientific (worldwide) info@AntecScientific.com www.AntecScientific.com T +31 172 268888

